Radiation Safety Program

Office of Environmental Health and Safety

2024
Cleveland State University
Radiation Safety Program

Purpose and Scope

The Cleveland State University (CSU) Radiation Safety Program is designed to provide information to applicable personnel and the public regarding the organization and management of radioactive materials on campus. It identifies regulations set forth by the Ohio Department of Health (ODH) and United States Nuclear Regulatory Commission (USNRC), and policies developed by CSU as safe, reasonable and enforceable. The manual is designed to conform closely to the Code of Federal Regulations (10CFR20) – "Standards for Protection Against Radiation" and similar regulations of the ODH as specified in the Ohio Administrative Code (OAC) Chapters 3701-38, 3701-39 and 3701-40. Copies of these regulations are on file in the Radiation Safety Office, located in the Department of Safety and Environmental Services, Plant Services Building, Room PS 233.

ALARA Goal

The main goal of the CSU Radiation Safety Program is to minimize exposure to radioactive materials, and to the radiation levels they may produce, to a level that is AS LOW AS REASONABLY ACHEIVABLE (ALARA). at Radv

Monitoring of all

Radiation Safety Committee

A Radiation Safety Committee is required by the University's Radioactive Materials license. The Radiation Safety Committee is an administrative committee whose members are appointed by the University

Personnel involved in Using Radioactive Materials

1. Authorized Users (Investigators)

   All individuals who intend on using radioisotope or any other form of ionizing radiation must provide a summary of their past experience and training in handling radioactive materials (See Appendix A). This summary must be submitted to the RSO for consideration by the University Radiation Safety Committee in order to authorize new investigators.

   Authorized users are responsible for ensuri

for. The permit holder is relieved of the responsibility for performing and recording monthly surveys and record-keeping duties. However, all other requirements in this Radiation Safety Manual must be maintained: refresher training is required annually, and all labeled rooms/areas are to be controlled as radiation use rooms/areas (i.e. no eating or drinking, etc.) The permit and laboratories are maintained so that work with radioactive materials may start with the next receipt of radioactive chemicals.

3. INACTIVE – when the laboratory will not be using radioactive materials for a period of more than 6 months, the Authorized User is strongly encouraged to make the permit inactive. Essentially this is a "temporary termination" where the laboratory is released for work not involving radiation, but the Authorized User may re-activate the permit by contacting EH&S. A laboratory will be ready for work involving radiation within 2 weeks of EH&S receiving the restart request. All the requirements for a termination must be met (no stock vials, waste is removed, and a termination survey). See Section D below. Once inactive, there are no restrictions on laboratories and no paperwork or retraining requirements. A permit may be inactive for 3 to 5 years after which radiation safety retraining may be required or the permit may be terminated.
4. TERMINATED – see Section D below. Depending on the amount of time elapsed since termination, a terminated authorization will need a new application submittal or if the time line is short it can be reauthorized.. Radiation safety retraining may also be required.

When an Authorized User is expected to be away for up to six weeks, the Authorized User must designate another Authorized User to act in his/her place. The Alternate Contact will be in charge of the laboratory work involving radiation on behalf of the Authorized User. The Alternate Contact will be contacted for authorization-related issues and will perform most of the Authorized User's responsibilities and functions, including minor requests for permit amendments such as adding/deleting users. The Authorized User will, upon return, countersign all those items which required signatures during the absence. When the Authorize User and the Alternate Contact are expected to be away at the same time, the permit holder must advise the Radiation Safety Officer of an alternate arrangement.

C: Authorization Termination

When authorization is expected to be terminated, the Authorized User must notify the Radiation Safety Officer as far in advance as possible. This is particularly important when the termination is due to the permit holder's leaving the University. The Authorized User will assist Radiation Safety Program personnel to ensure that:

1. All radioactive materials still in the possession of the authorized holder are properly transferred, or disposed of.
2. Personnel monitoring is discontinued, where applicable.
3. The laboratory area is free of radioactive contamination and radiation caution signs, labels and other postings are removed.
4. EH&S will perform an exit survey of the laboratory and equipment to ensure all spaces and items are free of contamination.

D: Authorized Assistants

to

determine whether or not they are sufficient. Authorized assistants are required to complete a Radiation Safety Training Course that includes passing a written examination.

Authorized Assistants must work under the supervision of an Authorized User. Authorized Assistants are responsible outlining and completing their assignments in as safe a manner as possible. They are to report any unsafe condition to the authorized user responsible for the area or to the RSO.

A. Students

ALL STUDENTS SHALL WORK UNDER THE DIRECT SUPERVISION OF AN AUTHORIZED USER.

Prior to students handling radioactive materials, the following procedures will be followed:

1. Specific procedures to be performed by students are to be submitted to the RSO for review and approval. The Radiation Safety Training Form and Radiation Worker Application shall be filled out and returned to the RSO or ARSO.

2. Students must be provided and with the radiation safety rules and policies for general laboratory safety. Each authorized worker shall review these rules and policies with each student worker.

3. The following aspects of the radiation safety program are to be reviewed and demonstrated to each student:

    a. Procedures for proper ordering and transfer of radioactive materials

    b. Techniques specific to the laboratory experiment being performed

    c. Techniques for monitoring physical facilities and personnel

    d. Procedures for proper disposal of radioactive waste

    e. Procedures for handling spills and emergencies

B. Ancillary Personnel

All ancillary personnel (campus police, building services and maintenance) who enter laboratories containing radioactive materials will receive training initially upon employment, and annually thereafter by the RSO. Training topics include:

1. Tsnidioanc(iv)]TJ a/MCID 84 -2(Tw 4.43 0v Tc -0.002 Tw 3.71 0 Td [(s)-9(p)2(ills)]TJ5)-4(d)]TJ c

E. Animal Use

The University's License from the Ohio Department of Health authorizes the use of radioactivity in animal studies. All research proposals identifying the need for working with radioactivity in

e. Place waste bags in a

All sealed sources shall be wiped checked for leakage every six months by the Radiation Safety Office. Samples shall be taken from the surface of the sealed source, or from surfaces of the container(s) in which the sealed source is mounted where contamination is likely to accumulate. If contained inside instrumentation or devices, the unit must be de-energized prior to performing the test.

Moistened filter paper is used to wipe the surface of the sealed source and counted by liquid scintillation. When testing the surface of the container(s) in which a sealed source is mounted, areas of approximately 100 cm^2 are wiped and counted by liquid scintillation. Removable contamination shall not exceed 0.005 uCi.

radioactive waste container. Also, put contaminated gloves and any other contaminated disposable material in the bag.

Appendices

A. Training and Experience of Users Form

B. Receipt of Radioactive Materials Form

C. Radioisotope Inventory Form

D. Radioactive Waste Log (Solids) Form

E. Sink Disposal of Radioactive Waste (Liquids) Form

F. Laboratory Inspection Form

G. Radioactivity Spill Report Form

H. Allowable Discharge to Sewer

I. Radiation Safety Laboratory Rules

J. Radiation Survey Equipment

K. Radioactive Materials

Appendix A

Application for Authorized Use of Radioactive Materials (RAM) Form

Cleveland State University
Radiation Safety Office
2121 Euclid Avenue, PS 233, Cleveland, Ohio 44115
(216) 687-9306 Fax (216) 687-9346

Application for Authorized Use of Radioactive Materials (RAM)

Applicant Information

Name_____
Department_____
Division_____
Position_____
CSU Office Location_____
Phone Extension_____

Highest degree:_____B.S._____M.S._____Ph.D._____M.D._____Other

If previously authorized for RAM uses on an NRC or State license, provide name and location of the institution:_____

Specific Education, Training anS (o)-14(cat)-62 4.65 02Tj -01255M0Td ( )Tj -0.01 DM3:

Appendix B

Receipt of Radioactive Materials Form

Shipment Receipt Form

SHIPMENT RECEIPT FORM

PURCHASE ORDER#_____

1. Authorized User:_____Survey Date_____Time_____
   Isotope and Activity:_____Surveyor:_____

2. Condition of Package:
   _____O.K._____Punctured_____Stain_____Wet
   _____Crushed _____Other
3. Smear Survey Required per NRC? _____Yes _____No
4. Radiation Label:
   _____White-I _____Yellow-II _____Yellow-III
5. Transportation Index:_____(0.25mR/hr at 3 feet)
6. Measure Radiation Levels:
   a. Package Surface:_____mR/hr
      200 mR/hr?_____Yes_____No (Yes requires notification)
   b. 3 feet from surface_____mR/hr
      10 mR/hr?_____Yes_____No (Yes requires notification)

7. Do Packing Slip and Vial Contents Agree?
   a. Radionuclide: <</MCID 60 >>BDC 78 -0 0 12 54 451.08 Tm 382.3Tj -0.81 Tf ( )Tj /TT1 1 Tf -0.002 Tc D

Appendix C

Radioactive Materials (RAM) Use Log Inventory Form

CLEVELAND STATE UNIVERSITY
RADIOACTIVE MATERIALS (RAM) ON HAND
(USE LOG)

Primary Investigator_____Room#_____

Date Received	Isotope	Inventory Date	Estimated RAD Level when Inventoried	Date Depeleted

Revised 1/07

Appendix D

Radioactive Waste Log
(Solids)

RADIOACTIVE WASTE LOG
(SOLIDS)

Location_____ Building_____

Date	Isotope	Type of Waste	Amount (uc)	Department	Disposed By	Phone

Revised 1/07

Appendix E

Sink Disposal of Radioactive Waste (Liquids)

SINK DISPOSAL OF RADIOACTIVE WASTE
(LIQUIDS)

User_____ Bldg_____ Lab Number_____

Dumping Limits: Not greater than 30 uci/day

Date	Radionuclide	Volume	Estimate of Activity Level	Initials

Appendix F

Radiation Laboratory
Inspection Form

Date:_____ Authorized User:_____

Laboratory Location:_____
(Building/Room

Appendix G

Radioactive Spill Report Form

RADIOACTIVE SPILL REPORT

Spill Occurred at____:____am/pm on / / in Bldg_____
Room_____

Instrument used to check for personnel contamination:

Meter Model _____ Meter S/N _____

Probe Model _____ Probe S/N _____

Personnel Present	Contamination Results

On the back of this sheet, indicate any personnel decontamination measures, additional monitoring, or care instituted.

Survey the spill area to identify hot spots, then begin decontamination. When finished, conduct a post 0 Td ()3(obe)]TJ

Appendix H

Allowable Release to Sewer of Radioactive Materials (RAM)

the sewer by the concentration of that radio nuclide listed in *OAC 3701:1-38-12*

Appendix I

Radiation Safety Laboratory Rules

Radiation Safety Laboratory Rules

1. All users of radioactive material must be approved by the Authorized User in this laboratory and first complete a radiation safety training program which includes passing a written exam. No student is permitted to handle radioactive material or to participate in an experiment otherwise

2. Laboratory coats and gloves must be worn when working with radioactive materials. Bench covering must be also used. Pipetting by mouth is forbidden.

3. Eating, drinking, chewing gum and application of cosmetics is expressly forbidden in the laboratory.

4. Open chemical operations should be normally avoided. The hood should be checked to establish the presence of airflow. In the event the amount of radioactive material will possibly exceed 100 uCi (microcuries) a hood must be used.

5. A radiation survey must be performed weekly in labs where radioactive material is used in amounts greater than 200 uCi (microcuries) at one time. Monthly surveys are required for amounts less than 200 uCi (microcuries) at one time. Radioactive material must be contained within the designated bench area or hood for radioactive experiments.

6. The movement of radioactive material outside the radiation laboratory is strictly forbidden. Transfer of radioactive materials between laboratories is permitted only with prior authorization by the Radiation Safety Officer.

7. It is required that all radioactive material be identified as to the radionuclide, amount of radioactivity (in uCi), and the date of assay. It is also required that the authorized user be identified on the label or storage area. Contaminated articles or glassware must be rinsed in the radioactive sink following the normal procedures for minimizing aerosol generation.

8. Film badges and/or ring badges must be worn if indicated at all times while the radioactive materials are being used. Film badges should remain in the pre-designated area when not worn and should never be taken off the premises. The Radiation Safety Office shall replace badges on a quarterly basis.

9. The Authorized User shall provide guidance and information on the proper disposal of radioactive waste in conjunction with the Radiation Safety Program. Contact the Radiation Safety Office for removal of radioactive waste from the lab.

10. All uses, transfers and disposition of radioactive material must be recorded in the radianui10. 3.7Tw 0.239 0sty

Appendix J

Radiation Survey Meters

Radiation Survey Equipment

Below is a list of equipment available for use in performing radiation safety surveys as part of the University's Radiati survey meters are calibrated on annual basis by a company certified to calibrate such meters. Calibration of all meters shall Radiation Safety Officer (RSO).

Manufacturer	**Model #**	Serial	Probe	User	Provided by CSU

Appendix K

Radioactive Materials Inter Laboratory Transfer Form

CLEVELAND STATE UNIVERSITY
Office of Environmental Health and Safety

RADIOACTIVE MATERIAL
INTER LABORATORY TRANSFER FORM

Authorized User_____ Date_____

Original Location (Building/Laboratory Number)_____

Isotope/Activity Transferred _____

Amount transferred_____

AuthorizedUser Receiving material_____

New Location (Building/LaboratoryNumber) _____

NOTE: *The radioactive wasteshould be recorded on the wastelog of the user* ***receiving*** *the RAM. This form should be put in the Note Book with the Use Logs.*

AuthorizationRSO /ARSO _____

Date_____

Appendix L

Radiation Safety Training Form

Environmental Health and Safety

Radiation Safety Training

Student Name: _____ Date:_____

CSU ID #:

Areas of Concentration:

Fundamentals of Radiation	Effects of Radiation
Radiation Exposure	Biological Effects
Dosimetry	Radiation Safety (ALARA)
Monitoring	Shielding

Cleveland State University Radiation Safety Procedures:

PLEASE POST

Penalties for Radiation Violations-Cleveland State University

Major Violations:

1. Release of radioisotope material into environment
2. Unreported spills of >100 microcuries of radioisotope that result in significant breach of containment, i.e. isotopes in non-radioactive trash, spills outside of PI's laboratory, etc.).
3. Unauthorized shipment of radioisotopes.
4. Ordering and/or receiving radioisotope not authorized by Radiation Safety Officer (RSO).

Any violation will result in an immediate suspension of radioisotope use and confiscation of all radioisotopes.

Moderate Violations:

1. Unauthorized transfer of radioisotopes between authorized labs. Transfer must be requested via INTER-LABORATORY TRANSFER FORM to Radiation Safety Officer (RSO). Both donor and recipient labs are responsible.
2. Unauthorized personnel using radioisotopes.
3. Food and/or drinks in the laboratory.
4. Evidence of eating and/or drinking in the laboratory.
5. Performing experiments not approved by user agreement.
6. Unreported spills of radioisotope greater than 100 microcuries.

One (1) violation within a twelve (12) month period will result in an official reprimand from the Safety Office indicating that the lab is placed on probation.
More than one (1) violation within a twelve (12) month period will result in an immediate suspension of radioisotope use and confiscation of all radioisotopes.

Minor Violations:

1. Unsecured radioisotopes (unlocked lab door or unlocked storage for radioisotopes)
2. Incomplete / inaccurate documentation including but not limited to the receipt of radioisotopes, post-experiment surveys, and disposal of radioactive waste (solid and liquid).
3. Performing experiments with inadequate personal protective equipment (PPE).
4. Performing experiments outside of areas designated for radioisotope use.
5. Use of Geiger counters or other safety devices that are not properly calibrated.

Two (2) violations within a twelve (12) month period will result in an official reprimand from the Radiation Safety Office indicating that the lab in violation is on probation.
More than two (2) violations within a twelve (12) month period will result in an immediate suspension of radioisotope use and confiscation of all radioisotopes.

Suspension: Any laboratory that is suspended for radioisotope use will be eligible to re-apply for radioisotope use no sooner than six (6) months from notification of violation. In order to be re-approved for radioisotope use, the PI and all lab members must undergo extensive training and testing to verify that further violations will not occur. The RSO and committee will make final decisions regarding the approval of radioisotope use for laboratories that have been suspended due to various safety infractions and violations.

Probation: Any laboratory that is placed under probation will remain under intensive safety reviews over the next twelve (12) months from notification of the violation. Any violation denoted within that twelve (12) month period will result in an immediate suspension of radioisotope use and confiscation of all radioisotopes.

Hearing: For moderate and major violations, the PI has the right to request a hearing by the Radiation Committee.

Appendix N



Environmental Health and Safety

Radiation Worker Information and Application Form

Employee: _____ Date:_____

Date of Birth: _____

CSU Identification Number: _____

Previous location where radioactive materials were used:_____

(exposure data will be requested)

Radioisotope(s) that will be used at CSU:

Short description of protocols for which radioisotopes will be used: